

Dispenser Selection Chart

Reagent	Dispensette® S	Dispensette® Organic	Reagent	Dispensette® S	Dispensette® Organic	Reagent	Dispensette® S	Dispensette® Organic
Acetaldehyde	+	+	Cyclohexane		+	Methylene chloride		+
Acetic acid (glacial), 100%	+	+	Cyclohexanone	+	+	Mineral oil (Engine oil)	+	+
Acetic acid, ≤ 96%	+	+	Cyclopentane		+	Monochloroacetic acid	+	+
Acetic anhydride		+	Decane	+	+	Nitric acid, ≤ 30%	+	+
Acetone	+	+	1-Decanol	+	+	Nitric acid, 30-70% */ **		+
Acetonitrile	+	+	Dibenzyl ether	+	+	Nitrobenzene	+	+
Acetophenone		+	Dichloroacetic acid		+	Oleic acid	+	+
Acetyl chloride		+	Dichlorobenzene	+	+	Oxalic acid	+	
Acetylacetone	+	+	Dichloroethane		+	n-Pentane		+
Acrylic acid	+	+	Dichloroethylene		+	Peracetic acid		+
Acrylonitrile	+	+	Dichloromethane		+	Perchloric acid	+	+
Adipic acid	+		Diesel oil (Heating oil), bp 250-350 °C		+	Perchloroethylene		+
Allyl alcohol	+	+	Diethanolamine	+	+	Petroleum, bp 180-220 °C		+
Aluminium chloride	+		Diethyl ether		+	Petroleum ether, bp 40-70 °C		+
Amino acids	+		Diethylamine	+	+	Phenol	+	+
Ammonia, ≤ 20%	+	+	1,2 Diethylbenzene	+	+	Phenylethanol	+	+
Ammonia, 20-30%		+	Diethylene glycol	+	+	Phenyldiazine	+	+
Ammonium chloride	+		Dimethyl sulfoxide (DMSO)	+	+	Phosphoric acid, ≤ 85%	+	+
Ammonium fluoride	+		Dimethylaniline	+		Phosphoric acid, 85% +	+	+
Ammonium sulfate	+		Dimethylformamide (DMF)	+	+	Sulfuric acid, 98%, 1:1		+
n-Amyl acetate	+	+	1,4 Dioxane		+	Piperidine	+	+
Amyl alcohol (Pentanol)	+	+	Diphenyl ether	+	+	Potassium chloride	+	
Amyl chloride (Chloropentane)	+		Essential oil		+	Potassium dichromate	+	
Aniline	+	+	Ethanol	+	+	Potassium hydroxide	+	
Barium chloride	+		Ethanolamine	+	+	Potassium permanganate	+	
Benzaldehyde	+	+	Ethyl acetate	+	+	Propionic acid	+	+
Benzene (Benzol)	+	+	Ethylbenzene		+	Propylene glycol (Propanediol)	+	+
Benzine (Petroleum benzin), bp 70-180 °C		+	Ethylene chloride		+	Pyridine	+	+
Benzoyl chloride	+	+	Fluoroacetic acid		+	Pyruvic acid	+	+
Benzyl alcohol	+	+	Formaldehyde, ≤ 40%	+		Salicylaldehyde	+	+
Benzylamine	+	+	Formamide	+	+	Scintillation fluid	+	+
Benzylchloride	+	+	Formic acid, ≤ 100%		+	Silver acetate	+	
Boric acid, ≤ 10%	+	+	Glycerol	+	+	Silver nitrate	+	
Bromobenzene	+	+	Glycol (Ethylene glycol)	+	+	Sodium acetate	+	
Bromonaphthalene	+	+	Glycolic acid, ≤ 50%	+		Sodium chloride	+	
Butanediol	+	+	Heating oil (Diesel oil), bp 250-350 °C		+	Sodium dichromate	+	
1-Butanol	+	+	Heptane		+	Sodium fluoride	+	
n-Butyl acetate	+	+	Hexane		+	Sodium hydroxide, ≤ 30%	+	
Butyl methyl ether	+	+	Hexanoic acid	+	+	Sodium hypochlorite	+	
Butylamine	+	+	Hexanol	+	+	Sulfuric acid, ≤ 98%	+	+
Butyric acid	+	+	Hydriodic acid, ≤ 57% **	+	+	Tartaric acid	+	
Calcium carbonate	+		Hydrobromic acid		+	Tetrachloroethylene		+
Calcium chloride	+		Hydrochloric acid, ≤ 20%	+	+	Tetrahydrofuran (THF) */ **		+
Calcium hydroxide	+		Hydrochloric acid, 20-37% **		+	Tetramethylammonium hydroxide	+	
Calcium hypochlorite	+		Hydrogen peroxide, ≤ 35%		+	Toluene		+
Carbon tetrachloride		+	Isoamyl alcohol	+	+	Trichloroacetic acid		+
Chloro naphthalene	+	+	Isobutanol	+	+	Trichlorobenzene		+
Chloroacetaldehyde, ≤ 45%	+	+	Isooctane		+	Trichloroethane		+
Chloroacetic acid	+	+	Isopropanol (2-Propanol)	+	+	Trichloroethylene		+
Chloroacetone	+	+	Isopropyl ether	+	+	Trichlorotrifluoro ethane		+
Chlorobenzene	+	+	Lactic acid	+		Triethanolamine	+	+
Chlorobutane	+	+	Methanol	+	+	Triethylene glycol	+	+
Chloroform		+	Methoxybenzene	+	+	Trifluoro ethane		+
Chlorosulfonic acid		+	Methyl benzoate	+	+	Trifluoroacetic acid (TFA)		+
Chromic acid, ≤ 50%	+	+	Methyl butyl ether	+	+	Turpentine		+
Chromosulfuric acid	+		Methyl ethyl ketone	+	+	Urea	+	
Copper sulfate	+		Methyl formate	+	+	Xylene		+
Cresol		+	Methyl propyl ketone	+	+	Zinc chloride, ≤ 10%	+	
Cumene (Isopropyl benzene)	+	+				Zinc sulfate, ≤ 10%	+	

The above recommendations reflect testing completed prior to publication. Always follow instructions in the operating manual of the instrument as well as the reagent manufacturer's specifications. In addition to these chemicals, a variety of organic and inorganic saline solutions (e.g., biological buffers), biological detergents and media for cell culture can be dispensed. Should you require information on chemicals not listed, please feel free to contact BRAND. Status as of: 0605/13

Note! For dispensing HF, we recommend the use of the Dispensette® S Trace Analysis bottle-top dispenser with platinum-iridium valve spring.



* use ETFE/PTFE bottle adapter

** use PTFE seal for valve block